



Pressure Gauges (90mm)

RODEM®

PROCESS EQUIPMENT

www.rodem.com 800-543-7312

The 90mm Anderson "Pharmaceutical" gauge has rapidly become the pressure gauge of choice for applications in the pharmaceutical and biotechnology industries. It is designed to meet the most stringent of specifications, including those for accuracy, surface finish, and temperature stability. Applications include process lines and vessels, filtration equipment, pumps, heat exchangers and other pressure dependent devices or processes. These gauges are particularly appropriate when the process line or equipment will be steam sterilized. The unfilled version can even be completely autoclaved while on the equipment.

A few of the unique features of this gauge include the

electropolished diaphragm and standard re-zero feature, along with a special polysulfone crystal that withstands autoclave temperatures as well as cleaning chemicals that readily etch other lenses. Each gauge is shipped with a certificate of calibration, your assurance that it has been calibrated to the most exacting tolerances.

A new available option provides increased dampening, without oil-filling the case.

For complete peace of mind in specifying sanitary pressure gauges that meet or exceed the CGMP's of the Pharmaceutical Industry, specify the Anderson "Pharmaceutical Series".

SPECIFICATIONS

PERFORMANCE

Calibrated Accuracy:

- ± 0.75% of full scale, from 10% - 90% of range (pressure ranges)
- ± 0.75% of full scale, from 25" vacuum to 90% of pressure range (compound ranges)

Repeatability:

- ± 0.25% of full scale

Hysteresis:

- ± 0.25% of full scale

Linearity:

- ± 0.25% of full scale

Operating Temperature Limits (process):

- 25 to 250°F (-3 to 124°C)

Ambient Temperature Limits:

- 40 to 120°F (4 to 49°C)

In-line Steam Sterilization Limits:

- Up to 300°F (149°C) for 1 hour

Autoclave Sterilization Limits:

- Up to 300°F (149°C) for 1 hour (unfilled case only)

Temperature Effect:

- ± 0.16% /10°F change in process or ambient temperature

Stability:

Within specified accuracy for 6 months under normal operating conditions. Calibration check recommended at 3 month intervals for gauges being steam sterilized or autoclaved

Over-range Capability:

- 10% of span

CONSTRUCTION/FINISH

Diaphragm:

- Welded 316L stainless steel, electropolished.
- Maximum R_a = 8 microinches (.2 microns)

Fitting:

- Welded 316L stainless steel polished (180 grit)

Case and Stem:

- Welded 304 stainless steel polished (180 grit)

Crystal:

- Polysulfone, .080" minimum thickness, corrosion resistant

Dial:

- 3.50 inch diameter (90 mm); pressure sensitive adhesive

Standards:

- Designed and manufactured to sound engineering practices in accordance with Article 3.3 of the PED 97/23/EC

Environmental Rating:

- NEMA 4X, IP-66

OPERATIONAL

Capillary (internal) Fill:

- 80% glycerine (food grade) / 20% USP de-ionized water

Re-zero Feature (standard):

- Tamper-resistant; ± 5% of span, non-interactive with span

Response Characteristics (Typical):

- Unfilled: 40%
- (Gauge dampens 40% of pulsation)
- Filled: 45%
- Mechanically Dampened: 95%

Complete Product Ordering Matrix



ANDERSON INSTRUMENT CO., INC.

156 Auriesville Road
Fultonville, NY 12072 USA

1-800-833-0081 • FAX 518-922-8997

ANDERSON INSTRUMENT CO. LP

400 Britannia Road East, Unit 1
Mississauga, Ontario L4Z 1X9 Canada

905-568-1440 • FAX 905-568-1652

NEGELE MESSTECHNIK GmbH
(A Division of Anderson)

Raiffeisenweg 7

D-87743 Egg a. d. Günz Germany

+49 (0) 8333/9204-0 • FAX +49 (0) 8333/9204-49

LATIN AMERICA AND ASIA

Maida International

516-676-3079 • FAX 516-676-3199

www.andinst.com

E B [] [] 0 1 [] [] [] [] [] []

	PRESSURE RANGE
029	30" Hg/0/30 psig (-1/0/2 Bar)
031	30" Hg/0/ 60 psig (-1/0/4 Bar)
032	30" Hg/0/100 psig (-1/0/7 Bar)
033	30" Hg/0/160 psig (-1/0/11 Bar)
034	30" Hg/0/300 psig (-1/0/20 Bar)
066	0-30 psig (0-2 Bar)
069	0-60 psig (0-4 Bar)
071	0-100 psig (0-7 Bar)
074	0-160 psig (0-11 Bar)
077	0-300 psig (0-20 Bar)
082	0-600 psig (0-40 Bar)
055	-1/0/2 Bar
056	-1/0/4 Bar
058	0/2.5 Bar
060	0/6 Bar
061	0/10 Bar
064	0/16 Bar
067	0/40 Bar
	DIAL NAME/LOGO
01	Anderson
	PROCESS CONNECTION
004	1-1/2" Tri-Clamp
005	2" Tri-Clamp
016	1-1/2" C.B. "I" Line (male)
017	2" C.B. "I" Line (male)
123	Anderson "CPM" Flushmount
	MOUNTING
1	Bottom
4	Back
5	Left
6	Right
	CASE FILL
0	Unfilled Case
2	Mechanical Dampening
	CAPILLARY FILL
1	Mineral Oil
4	80/20 Glycerin/H ₂ O (Standard)
5	Propylene Glycol (Neobee M-20)
	CRYSTAL
3*	Polysulfone

* Polysulfone with unfilled case enables gauge to withstand autoclave cycles